



Limited liability company
Gazprom Pererabotka
(Gazprom Pererabotka, LLC)

**The branch
Condensate Stabilization Plant
named after V.S. Chernomyrdin
(Surgut CSP)**

**ATTN: O.N. Ekaterinina
Director General
of SKB EP, LLC**

p/b 1233, uts Solnechnyi, Surgut district,
Khanty-Mansi Autonomous Okrug–Yugra, Russia, 628452
Tel.: +7 (3462) 95-45-19, +7 (725) 5-45-19; Fax: +7 (3462) 28-41-71

09.11.2020 № 302/79/01284

на № _____ от _____

Review to the use of MIKO-8 milli-ohmmeter

Dear Olga Nikolaevna!

Our plant has been using MIKO-8 milliohmmeter since 2016. The device is built into a good strong case that is convenient for transportation. During the time of use, with the help of MIKO-8 were diagnosed more than 100 units of electrical equipment (transformers, electric motors of various capacities and other equipment). One of the main advantages for our staff was the ability of automatic recalculations of deviations with indication of parameters and their saving into the device memory. The second advantage is the possibility of further processing of measurement results on a PC using specialized software. Also, among the advantages we can highlight the measurement speed, intuitive operation of the device, and its small dimensions.

Among the disadvantages, we note the use of non-standard connectors on the test cables and the inability to use third-party cables, the lack of auto power selection of the device.

MIKO-8 copes well with the assigned tasks. We can recommend it to other companies.

We allow you to post this review on the website and to use it for advertising purposes.

Chief Engineer

S.L. Ivanov



79 50923638 8