



ПУБЛИЧНОЕ АКЦИОНЕРНОЕ ОБЩЕСТВО
«ЭЛЕКТРОВЫПРЯМИТЕЛЬ»
PUBLIC JOINT-STOCK COMPANY "ELECTROVIPRYAMITEL"

РОССИЯ, МОРДОВИЯ, г. САРАНСК, 430001, ул. Пролетарская, 126
RUSSIA, MORDOVIA, SARANSK, 430001, Proletarskaya Str, 126

TESTING CENTER

Tel.: 8 (8342) 47-17-44; E-mail: elvpr.ic@yandex.ru

№ 196
«20» 10 2020г.

ATTN: Director General
of SKB EP, LLC
O.N. Ekaterinina

Review to MIKO-10 micro-ohmmeter

Small-sized wrist micro-ohmmeter MIKO-10 manufactured by SKB EP, LLC was purchased by the Testing Center in July 2019, in order to replace E6-18 milliohmmeter, that no longer meet the required accuracy characteristics or test conditions.

MIKO-10 have been used during standard, periodic, certification tests of products manufactured by Elektrovypryamitel, PJSC, for:

- the electrical resistance grounding measurement of current-carrying parts accessible to touch;
- measurement of electrical resistance to direct current.

During the operation the following positive aspects were revealed:

- the device is compact, lightweight and has several ways of fixing and while operating hands remain free, that is necessary for measurements record;
- the device is simple-to-operate, also it is possible to save the results to a computer;
- the device is listed in the State Register of the Russian Federation;
- good operation independence.

There is one disadvantage revealed. During operation, the device switches off automatically. We have to turn it on with the power button, while the charge level is normal, and can't cause device's shutdown.

Head of the Testing Centre of
Elektrovypryamitel, PJSC

A.I. Krainov