

REPORT

31 January 2023

Irkutsk

During the use of ZU-2A charger when working on Taishet grid in the period from 17.11.2022 to 26.11.2022 a significant advantage over the standard complete set was revealed in terms of measurements duration without recharging. The operating time with MIKO-2.3 instrument amounted 8 hours without recharging. We also would like to note quick charging time. One of the disadvantages is a short connector cable of ZU-2A charger to MIKO-2.3. It would be much more convenient, if a connection plug was at a right angle. In this case, a cable break near the plug would be excluded.

Deputy Head of the Electrotechnical Laboratory



S.V. Gromyshev