



MOSKABELMET

group of company

Moskabel Plant, LLC

2 2nd Kabelnaya St.,
Moscow, Russia, 111024
Tel.: +7 (495) 673-52-81
E-mail: sale@ck.mkm.ru

ATTN: O.N. Ekaterinina
Director General
of SKB EP, LLC

Review about the use of the device

Dear Olga Nikolaevna!

Currently, in the process of conductive cores production at Moskabel Plant, LLC, MIKO-10 microohmmeter is used for post-operative control during the adjustment of the production of TPJ of 1-2 flexibility class according to GOST 22483-2012 to determine the resistance of TPJ. The compact size of the device allows it to be used in conditions of limited space, at the same time, the readings of the device differ in one area within a series of consecutive measurements. At the moment, the use of MIKO-10 enables reducing the filling length by 1-3 meters.

As of now, we can recommend the use of MIKO-10 with great caution in the conditions of manufacturers of cable and wire products.

Deputy Director
for Technology and Development

D.I. Solomatin